

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006681.D
 Acq On : 20 Dec 2018 17:34
 Operator : JC/MD
 Sample : J6484-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 976-MW-20(22)

Quant Time: Dec 21 04:22:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	617808	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	979870	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	901852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	425228	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	341748	51.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.66%	
35) Dibromofluoromethane	5.50	113	289074	46.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.30%	
50) Toluene-d8	8.71	98	1164449	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
62) 4-Bromofluorobenzene	11.13	95	413734	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
Target Compounds						
16) Acetone	2.45	43	5787	1.892	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	3758	0.557	ug/l	94
44) Trichloroethene	7.21	130	11593	1.553	ug/l	89
64) Tetrachloroethene	9.34	164	6524	0.914	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006681.D
Acq On : 20 Dec 2018 17:34
Operator : JC/MD
Sample : J6484-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
976-MW-20(22)

Quant Time: Dec 21 04:22:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

