

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006686.D
 Acq On : 20 Dec 2018 19:47
 Operator : JC/MD
 Sample : J6484-13 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 973-MW-17(15.5)

Quant Time: Dec 21 04:35:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	624258	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	989202	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	922019	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	438438	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	343885	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
35) Dibromofluoromethane	5.50	113	299148	47.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.64%	
50) Toluene-d8	8.71	98	1184586	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
62) 4-Bromofluorobenzene	11.13	95	428713	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	19954	2.926	ug/l	83
44) Trichloroethene	7.21	130	20989	2.785	ug/l	98
64) Tetrachloroethene	9.34	164	748833	102.566	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006686.D
Acq On : 20 Dec 2018 19:47
Operator : JC/MD
Sample : J6484-13 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
973-MW-17(15.5)

Quant Time: Dec 21 04:35:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

