

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006693.D
 Acq On : 20 Dec 2018 23:47
 Operator : JC/MD
 Sample : J6514-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-TB-2018-2

Quant Time: Dec 21 10:23:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	655585	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1012538	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	924989	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	437133	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	348069	49.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.52%	
35) Dibromofluoromethane	5.51	113	308001	48.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.20%	
50) Toluene-d8	8.71	98	1200215	49.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.44%	
62) 4-Bromofluorobenzene	11.13	95	426567	48.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.50%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006693.D
Acq On : 20 Dec 2018 23:47
Operator : JC/MD
Sample : J6514-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-TB-2018-2

Quant Time: Dec 21 10:23:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

