

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006695.D
 Acq On : 21 Dec 2018 00:41
 Operator : JC/MD
 Sample : J6514-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0494

Quant Time: Dec 21 04:52:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	619770	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	981941	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	909773	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	425701	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	337055	50.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.92%	
35) Dibromofluoromethane	5.50	113	294988	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	8.71	98	1163944	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
62) 4-Bromofluorobenzene	11.13	95	419356	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	100638	15.775	ug/l	99
16) Acetone	2.45	43	24625	8.024	ug/l	100
21) trans-1,2-Dichloroethene	3.17	96	12715	2.072	ug/l	92
24) 1,1-Dichloroethane	3.70	63	4445	0.437	ug/l #	82
27) cis-1,2-Dichloroethene	4.60	96	469252	69.319	ug/l	93
44) Trichloroethene	7.21	130	1930873	258.062	ug/l	99
52) Toluene	8.78	92	5876	0.368	ug/l	86
64) Tetrachloroethene	9.34	164	1913	0.266	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006695.D
Acq On : 21 Dec 2018 00:41
Operator : JC/MD
Sample : J6514-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0494

Quant Time: Dec 21 04:52:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

