

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006696.D
 Acq On : 21 Dec 2018 01:07
 Operator : JC/MD
 Sample : J6514-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0495

Quant Time: Dec 21 04:54:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	637944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1016039	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	925047	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	436722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	348896	50.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.48%	
35) Dibromofluoromethane	5.50	113	304044	47.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.64%	
50) Toluene-d8	8.71	98	1204023	49.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.40%	
62) 4-Bromofluorobenzene	11.13	95	420224	47.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.74%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	110939	16.894	ug/l	99
16) Acetone	2.45	43	22255	7.045	ug/l	99
21) trans-1,2-Dichloroethene	3.17	96	11072	1.753	ug/l	95
27) cis-1,2-Dichloroethene	4.60	96	579840	83.216	ug/l	91
44) Trichloroethene	7.21	130	2129216	275.020	ug/l	99
52) Toluene	8.79	92	6449	0.390	ug/l	91
64) Tetrachloroethene	9.34	164	13373	1.826	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006696.D
Acq On : 21 Dec 2018 01:07
Operator : JC/MD
Sample : J6514-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0495

Quant Time: Dec 21 04:54:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

