

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122018\
 Data File : VX006707.D
 Acq On : 21 Dec 2018 06:01
 Operator : JC/MD
 Sample : J6517-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-MW-1-9.0-121918

Quant Time: Dec 21 10:52:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	656003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1009463	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	906372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	428227	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	343460	48.58	ug/l	0.00
Spiked Amount			Recovery	=	97.16%	
35) Dibromofluoromethane	5.51	113	296250	46.41	ug/l	0.00
Spiked Amount			Recovery	=	92.82%	
50) Toluene-d8	8.71	98	1182698	49.14	ug/l	0.00
Spiked Amount			Recovery	=	98.28%	
62) 4-Bromofluorobenzene	11.13	95	420439	47.70	ug/l	0.00
Spiked Amount			Recovery	=	95.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	16371	5.040	ug/l	92
31) Cyclohexane	5.58	56	7264	0.742	ug/l #	80
39) Methylcyclohexane	7.46	83	6199	0.563	ug/l	91
40) Benzene	6.15	78	9186	0.361	ug/l #	89
44) Trichloroethene	7.21	130	2857	0.371	ug/l	76
52) Toluene	8.79	92	6218	0.379	ug/l	93
68) m/p-Xylenes	10.36	106	10699	0.865	ug/l	92
69) o-Xylene	10.70	106	3594	0.294	ug/l	97
73) Isopropylbenzene	11.02	105	41407	1.413	ug/l	99
78) n-propylbenzene	11.36	91	68656	2.128	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	8387	0.330	ug/l	96
85) sec-Butylbenzene	11.94	105	10611	0.358	ug/l #	72
95) Naphthalene	13.83	128	7944	0.253	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122018\
Data File : VX006707.D
Acq On : 21 Dec 2018 06:01
Operator : JC/MD
Sample : J6517-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-MW-1-9.0-121918

Quant Time: Dec 21 10:52:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

