

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX122018\
 Data File : VX006662.D
 Acq On : 20 Dec 2018 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 03:44:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	118	0.00
2 T	Dichlorodifluoromethane	0.457	0.430	5.9	116	0.00
3 P	Chloromethane	0.493	0.462	6.3	116	0.00
4 C	Vinyl Chloride	0.515	0.484	6.0#	116	0.00
5 T	Bromomethane	0.560	0.497	11.3	120	0.00
6 T	Chloroethane	0.356	0.311	12.6	115	0.00
7 T	Trichlorofluoromethane	0.802	0.766	4.5	116	0.00
8 T	Diethyl Ether	0.319	0.295	7.5	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.471	0.470	0.2	122	0.00
10 T	Methyl Iodide	0.674	0.705	-4.6	116	0.00
11 T	Tert butyl alcohol	0.120	0.110	8.3	110	0.00
12 CM	1,1-Dichloroethene	0.447	0.419	6.3#	117	0.00
13 T	Acrolein	0.081	0.065	19.8	99	0.00
14 T	Allyl chloride	0.752	0.721	4.1	116	0.00
15 T	Acrylonitrile	0.267	0.252	5.6	113	0.00
16 T	Acetone	0.248	0.250	-0.8	125	0.00
17 T	Carbon Disulfide	1.182	1.073	9.2	114	0.00
18 T	Methyl Acetate	0.728	0.694	4.7	116	0.00
19 T	Methyl tert-butyl Ether	1.483	1.448	2.4	117	0.00
20 T	Methylene Chloride	0.512	0.469	8.4	117	0.00
21 T	trans-1,2-Dichloroethene	0.495	0.453	8.5	117	0.00
22 T	Diisopropyl ether	1.376	1.406	-2.2	122	0.00
23 T	Vinyl Acetate	1.125	1.182	-5.1	125	0.00
24 P	1,1-Dichloroethane	0.821	0.796	3.0	122	0.00
25 T	2-Butanone	0.348	0.340	2.3	118	0.00
26 T	2,2-Dichloropropane	0.634	0.624	1.6	119	0.00
27 T	cis-1,2-Dichloroethene	0.546	0.518	5.1	115	0.00
28 T	Bromochloromethane	0.378	0.345	8.7	115	0.00
29 T	Tetrahydrofuran	0.229	0.217	5.2	114	0.00
30 C	Chloroform	0.850	0.826	2.8#	118	0.00
31 T	Cyclohexane	0.746	0.725	2.8	118	0.00
32 T	1,1,1-Trichloroethane	0.718	0.712	0.8	116	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.500	7.2	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	113	0.00
35 S	Dibromofluoromethane	0.316	0.313	0.9	112	0.00
36 T	1,1-Dichloropropene	0.414	0.420	-1.4	118	0.00
37 T	Ethyl Acetate	0.431	0.430	0.2	118	0.00
38 T	Carbon Tetrachloride	0.402	0.417	-3.7	118	0.00
39 T	Methylcyclohexane	0.545	0.557	-2.2	121	0.00
40 TM	Benzene	1.259	1.254	0.4	116	0.00
41 T	Methacrylonitrile	0.229	0.237	-3.5	114	0.00
42 TM	1,2-Dichloroethane	0.429	0.421	1.9	118	0.00
43 T	Isopropyl Acetate	0.656	0.693	-5.6	119	0.00
44 TM	Trichloroethene	0.381	0.377	1.0	119	0.00
45 C	1,2-Dichloropropane	0.306	0.312	-2.0#	122	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX122018\
 Data File : VX006662.D
 Acq On : 20 Dec 2018 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 03:44:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.225	0.222	1.3	115	0.00
47 T	Bromodichloromethane	0.374	0.396	-5.9	119	0.00
48 T	Methyl methacrylate	0.334	0.357	-6.9	118	0.00
49 T	1,4-Dioxane	0.009	0.009	0.0	112	0.00
50 S	Toluene-d8	1.192	1.188	0.3	113	0.00
51 T	4-Methyl-2-Pentanone	0.417	0.438	-5.0	117	0.00
52 CM	Toluene	0.813	0.843	-3.7#	119	0.00
53 T	t-1,3-Dichloropropene	0.401	0.450	-12.2	120	0.00
54 T	cis-1,3-Dichloropropene	0.440	0.492	-11.8	119	0.00
55 T	1,1,2-Trichloroethane	0.333	0.343	-3.0	120	0.00
56 T	Ethyl methacrylate	0.470	0.504	-7.2	116	0.00
57 T	1,3-Dichloropropane	0.536	0.539	-0.6	118	0.00
58 T	2-Chloroethyl Vinyl ether	0.250	0.249	0.4	114	0.00
59 T	2-Hexanone	0.319	0.342	-7.2	119	0.00
60 T	Dibromochloromethane	0.307	0.352	-14.7	122	0.00
61 T	1,2-Dibromoethane	0.354	0.372	-5.1	118	0.00
62 S	4-Bromofluorobenzene	0.437	0.448	-2.5	114	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	116	0.00
64 T	Tetrachloroethene	0.396	0.389	1.8	121	0.00
65 PM	Chlorobenzene	1.058	1.048	0.9	119	0.00
66 T	1,1,1,2-Tetrachloroethane	0.330	0.357	-8.2	121	0.00
67 C	Ethyl Benzene	1.729	1.739	-0.6#	120	0.00
68 T	m/p-Xylenes	0.682	0.696	-2.1	120	0.00
69 T	o-Xylene	0.674	0.681	-1.0	118	0.00
70 T	Styrene	1.066	1.111	-4.2	118	0.00
71 P	Bromoform	0.251	0.276	-10.0	120	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
73 T	Isopropylbenzene	3.421	3.469	-1.4	120	0.00
74 T	N-amyl acetate	1.224	1.294	-5.7	120	0.00
75 P	1,1,2,2-Tetrachloroethane	1.017	1.012	0.5	116	0.00
76 T	1,2,3-Trichloropropane	0.929	0.953	-2.6	115	0.00
77 T	Bromobenzene	0.925	0.926	-0.1	117	0.00
78 T	n-propylbenzene	3.767	3.934	-4.4	122	0.00
79 T	2-Chlorotoluene	2.283	2.277	0.3	120	0.00
80 T	1,3,5-Trimethylbenzene	2.817	2.870	-1.9	119	0.00
81 T	trans-1,4-Dichloro-2-butene	0.259	0.287	-10.8	122	0.00
82 T	4-Chlorotoluene	2.618	2.636	-0.7	120	0.00
83 T	tert-Butylbenzene	2.820	2.926	-3.8	119	0.00
84 T	1,2,4-Trimethylbenzene	2.963	2.960	0.1	120	0.00
85 T	sec-Butylbenzene	3.457	3.558	-2.9	120	0.00
86 T	p-Isopropyltoluene	3.039	3.215	-5.8	122	0.00
87 T	1,3-Dichlorobenzene	1.667	1.683	-1.0	119	0.00
88 T	1,4-Dichlorobenzene	1.709	1.686	1.3	119	0.00
89 T	n-Butylbenzene	2.586	2.810	-8.7	126	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX122018\
Data File : VX006662.D
Acq On : 20 Dec 2018 09:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 03:44:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.414	0.471	-13.8	126	0.00
91 T	1,2-Dichlorobenzene	1.667	1.623	2.6	117	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.210	0.213	-1.4	115	0.00
93 T	1,2,4-Trichlorobenzene	1.150	1.207	-5.0	123	0.00
94 T	Hexachlorobutadiene	0.542	0.563	-3.9	127	0.00
95 T	Naphthalene	3.671	3.644	0.7	117	0.00
96 T	1,2,3-Trichlorobenzene	1.187	1.197	-0.8	123	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6