

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122019\
 Data File : VX014167.D
 Acq On : 20 Dec 2019 12:33
 Operator : JC/SP
 Sample : K6391-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 20 23:43:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	589391	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	896958	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	783356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	347001	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	321319	48.10	ug/l	0.00
Spiked Amount						
			Recovery	=	96.20%	
35) Dibromofluoromethane	5.49	113	264348	48.48	ug/l	0.00
Spiked Amount						
			Recovery	=	96.96%	
50) Toluene-d8	8.71	98	1052278	49.57	ug/l	0.00
Spiked Amount						
			Recovery	=	99.14%	
62) 4-Bromofluorobenzene	11.13	95	348737	44.94	ug/l	0.00
Spiked Amount						
			Recovery	=	89.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122019\
Data File : VX014167.D
Acq On : 20 Dec 2019 12:33
Operator : JC/SP
Sample : K6391-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Dec 20 23:43:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

