

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122019\
 Data File : VX014181.D
 Acq On : 20 Dec 2019 17:59
 Operator : JC/SP
 Sample : K6372-12DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3DL

Manual Integrations
 APPROVED

apatel
 12/23/2019 10:25:14 AM

Quant Time: Jan 04 02:29:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	558376	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	853450	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	759476	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	350722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	306325	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	
35) Dibromofluoromethane	5.49	113	251981	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
50) Toluene-d8	8.71	98	1008834	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	11.13	95	340802	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.45	83	12709	1.314	ug/l #	85
40) Benzene	6.14	78	130968	5.436	ug/l	98
52) Toluene	8.78	92	31877	2.093	ug/l	100
67) Ethyl Benzene	10.25	91	869377	31.315	ug/l	99
68) m/p-Xylenes	10.36	106	236355	22.135	ug/l	100
69) o-Xylene	10.69	106	40950	3.910	ug/l	95
73) Isopropylbenzene	11.01	105	54929	2.224	ug/l	94
78) n-propylbenzene	11.35	91	131929	4.763	ug/l	99
80) 1,3,5-Trimethylbenzene	11.50	105	38337	1.845	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	758168	36.429	ug/l	99
85) sec-Butylbenzene	11.94	105	16584m	0.695	ug/l	
86) p-Isopropyltoluene	12.06	119	10045	0.460	ug/l	84
89) n-Butylbenzene	12.38	91	14385m	0.772	ug/l	
95) Naphthalene	13.83	128	368097	16.081	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122019\
Data File : VX014181.D
Acq On : 20 Dec 2019 17:59
Operator : JC/SP
Sample : K6372-12DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3DL

Manual Integrations
APPROVED

apatel
12/23/2019 10:25:14 AM

Quant Time: Jan 04 02:29:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

