

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX122019\
 Data File : VX014163.D
 Acq On : 20 Dec 2019 10:28
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 20 23:16:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.447	0.416	6.9	93	0.00
3 P	Chloromethane	0.600	0.540	10.0	89	0.00
4 C	Vinyl Chloride	0.633	0.582	8.1#	93	0.00
5 T	Bromomethane	0.450	0.402	10.7	93	0.00
6 T	Chloroethane	0.391	0.370	5.4	96	0.00
7 T	Trichlorofluoromethane	0.784	0.766	2.3	98	0.00
8 T	Diethyl Ether	0.368	0.353	4.1	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.473	0.470	0.6	101	0.00
10 T	Methyl Iodide	0.575	0.519	9.7	82	0.00
11 T	Tert butyl alcohol	0.122	0.142	-16.4	122	-0.01
12 CM	1,1-Dichloroethene	0.481	0.458	4.8#	96	0.00
13 T	Acrolein	0.070	0.083	-18.6	124	0.00
14 T	Allyl chloride	0.859	0.851	0.9	98	0.00
15 T	Acrylonitrile	0.286	0.299	-4.5	106	0.00
16 T	Acetone	0.368	0.298	19.0	78	0.00
17 T	Carbon Disulfide	1.418	1.216	14.2	91	0.00
18 T	Methyl Acetate	0.729	0.764	-4.8	107	0.00
19 T	Methyl tert-butyl Ether	1.580	1.643	-4.0	102	0.00
20 T	Methylene Chloride	0.579	0.534	7.8	98	0.00
21 T	trans-1,2-Dichloroethene	0.530	0.503	5.1	97	0.00
22 T	Diisopropyl ether	1.711	1.729	-1.1	100	0.00
23 T	Vinyl Acetate	1.397	1.532	-9.7	107	0.00
24 P	1,1-Dichloroethane	0.952	0.919	3.5	98	0.00
25 T	2-Butanone	0.454	0.448	1.3	96	0.00
26 T	2,2-Dichloropropane	0.739	0.748	-1.2	100	0.00
27 T	cis-1,2-Dichloroethene	0.599	0.590	1.5	100	0.00
28 T	Bromochloromethane	0.332	0.356	-7.2	92	0.00
29 T	Tetrahydrofuran	0.254	0.278	-9.4	109	0.00
30 C	Chloroform	0.903	0.902	0.1#	99	0.00
31 T	Cyclohexane	0.847	0.835	1.4	96	0.00
32 T	1,1,1-Trichloroethane	0.769	0.754	2.0	98	0.00
33 S	1,2-Dichloroethane-d4	0.567	0.549	3.2	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.304	0.302	0.7	95	0.00
36 T	1,1-Dichloropropene	0.459	0.453	1.3	97	0.00
37 T	Ethyl Acetate	0.507	0.560	-10.5	105	0.00
38 T	Carbon Tetrachloride	0.418	0.435	-4.1	98	0.00
39 T	Methylcyclohexane	0.566	0.580	-2.5	100	0.00
40 TM	Benzene	1.412	1.392	1.4	97	0.00
41 T	Methacrylonitrile	0.281	0.303	-7.8	106	-0.01
42 TM	1,2-Dichloroethane	0.479	0.491	-2.5	101	0.00
43 T	Isopropyl Acetate	0.838	0.910	-8.6	105	0.00
44 TM	Trichloroethene	0.390	0.383	1.8	100	0.00
45 C	1,2-Dichloropropane	0.362	0.369	-1.9#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX122019\
 Data File : VX014163.D
 Acq On : 20 Dec 2019 10:28
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 20 23:16:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 17 03:01:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.239	0.244	-2.1	102	0.00
47 T	Bromodichloromethane	0.451	0.480	-6.4	102	0.00
48 T	Methyl methacrylate	0.411	0.444	-8.0	103	0.00
49 T	1,4-Dioxane	0.008	0.010	-25.0#	124	0.00
50 S	Toluene-d8	1.183	1.168	1.3	92	0.00
51 T	4-Methyl-2-Pentanone	0.527	0.570	-8.2	104	0.00
52 CM	Toluene	0.892	0.888	0.4#	98	0.00
53 T	t-1,3-Dichloropropene	0.494	0.539	-9.1	100	0.00
54 T	cis-1,3-Dichloropropene	0.550	0.603	-9.6	101	0.00
55 T	1,1,2-Trichloroethane	0.358	0.371	-3.6	101	0.00
56 T	Ethyl methacrylate	0.562	0.610	-8.5	101	0.00
57 T	1,3-Dichloropropane	0.602	0.620	-3.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.213	0.240	-12.7	102	0.00
59 T	2-Hexanone	0.424	0.447	-5.4	99	0.00
60 T	Dibromochloromethane	0.355	0.393	-10.7	102	0.00
61 T	1,2-Dibromoethane	0.366	0.389	-6.3	101	0.00
62 S	4-Bromofluorobenzene	0.433	0.432	0.2	94	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.442	0.411	7.0	91	0.00
65 PM	Chlorobenzene	1.063	1.084	-2.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.397	-4.5	100	0.00
67 C	Ethyl Benzene	1.828	1.894	-3.6#	98	0.00
68 T	m/p-Xylenes	0.703	0.726	-3.3	98	0.00
69 T	o-Xylene	0.689	0.701	-1.7	97	0.00
70 T	Styrene	1.166	1.221	-4.7	97	0.00
71 P	Bromoform	0.305	0.341	-11.8	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.520	3.597	-2.2	98	0.00
74 T	N-amyl acetate	1.650	1.749	-6.0	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.220	1.254	-2.8	100	0.00
76 T	1,2,3-Trichloropropane	1.085	1.028	5.3	89	0.00
77 T	Bromobenzene	0.969	0.947	2.3	100	0.00
78 T	n-propylbenzene	3.949	4.110	-4.1	99	0.00
79 T	2-Chlorotoluene	2.400	2.428	-1.2	98	0.00
80 T	1,3,5-Trimethylbenzene	2.963	3.029	-2.2	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.385	0.431	-11.9	102	0.00
82 T	4-Chlorotoluene	2.784	2.820	-1.3	100	0.00
83 T	tert-Butylbenzene	2.883	3.029	-5.1	103	0.00
84 T	1,2,4-Trimethylbenzene	2.967	3.073	-3.6	99	0.00
85 T	sec-Butylbenzene	3.404	3.545	-4.1	100	0.00
86 T	p-Isopropyltoluene	3.114	3.309	-6.3	101	0.00
87 T	1,3-Dichlorobenzene	1.698	1.689	0.5	100	0.00
88 T	1,4-Dichlorobenzene	1.726	1.709	1.0	102	0.00
89 T	n-Butylbenzene	2.657	2.888	-8.7	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX122019\
Data File : VX014163.D
Acq On : 20 Dec 2019 10:28
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 20 23:16:54 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 17 03:01:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.559	0.584	-4.5	99	0.00
91 T	1,2-Dichlorobenzene	1.709	1.683	1.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.270	0.272	-0.7	103	0.00
93 T	1,2,4-Trichlorobenzene	1.111	1.204	-8.4	106	0.00
94 T	Hexachlorobutadiene	0.534	0.571	-6.9	107	0.00
95 T	Naphthalene	3.263	3.680	-12.8	105	0.00
96 T	1,2,3-Trichlorobenzene	1.096	1.178	-7.5	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6