

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122021\
 Data File : VX025936.D
 Acq On : 20 Dec 2021 23:54
 Operator : JC/MD
 Sample : M5150-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GBBX4

Quant Time: Dec 21 02:06:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML122021WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Dec 21 01:32:07 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	145423	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	143748	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	61273	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	55824	47.615	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	95.220%	
7) Chloroethane-d5	1.666	69	49940	51.010	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	102.020%	
11) 1,1-Dichloroethene-d2	2.306	63	79644	39.112	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	78.220%	
21) 2-Butanone-d5	4.458	46	94808	112.095	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	112.090%	
24) Chloroform-d	5.062	84	111253	49.503	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	99.000%	
26) 1,2-Dichloroethane-d4	5.964	65	76823	52.411	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	104.820%	
32) Benzene-d6	5.982	84	225209	51.712	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	103.420%	
36) 1,2-Dichloropropane-d6	7.311	67	67852	51.153	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	102.300%	
41) Toluene-d8	8.652	98	207765	50.244	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	100.480%	
43) trans-1,3-Dichloroprop...	8.951	79	25585	43.685	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.380%	
47) 2-Hexanone-d5	9.384	63	64151	107.586	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	107.590%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	94797	50.447	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	100.900%	
66) 1,2-Dichlorobenzene-d4	12.323	152	68450	52.999	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.000%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	1922	2.111	ug/L #	1
13) Acetone	2.379	43	3606	5.329	ug/L #	42
18) Methyl tert-butyl Ether	3.117	73	3671	1.067	ug/L	98
20) cis-1,2-Dichloroethene	4.501	96	8240	6.767	ug/L	100
34) Trichloroethene	7.128	95	1126395	957.433	ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122021\
Data File : VX025936.D
Acq On : 20 Dec 2021 23:54
Operator : JC/MD
Sample : M5150-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GBBX4

Quant Time: Dec 21 02:06:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM122021WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Dec 21 01:32:07 2021
Response via : Initial Calibration

