

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
 Data File : VX033504.D
 Acq On : 20 Dec 2022 13:45
 Operator : JC/MD
 Sample : N6149-03 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MAH-1

Quant Time: Dec 21 04:14:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	109080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	173974	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	163516	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85720	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	32506	46.451	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.900%
35) Dibromofluoromethane	5.385	113	37348	47.729	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	8.647	98	78096	47.708	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.420%
62) 4-Bromofluorobenzene	11.079	95	65321	52.188	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.380%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	4054	18.949	ug/l #	91
16) Acetone	2.386	43	7524	14.530	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
Data File : VX033504.D
Acq On : 20 Dec 2022 13:45
Operator : JC/MD
Sample : N6149-03 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MAH-1

Quant Time: Dec 21 04:14:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

