

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
 Data File : VX033508.D
 Acq On : 20 Dec 2022 15:17
 Operator : JC/MD
 Sample : N6149-06 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WANA-1

Quant Time: Dec 21 04:15:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189249	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	177432	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38603	51.360	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 102.720%		
35) Dibromofluoromethane	5.391	113	44983	52.846	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 105.700%		
50) Toluene-d8	8.647	98	92170	51.761	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 103.520%		
62) 4-Bromofluorobenzene	11.079	95	75897	55.744	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 111.480%		
Target Compounds						
					Qvalue	
8) Diethyl Ether	2.136	74	3176	4.335	ug/l #	31
11) Tert butyl alcohol	2.965	59	14834	64.554	ug/l #	90
16) Acetone	2.386	43	25393	45.657	ug/l	99
18) Methyl Acetate	2.703	43	2891	1.796	ug/l	94
19) Methyl tert-butyl Ether	3.117	73	5232	1.422	ug/l #	61
25) 2-Butanone	4.568	43	13972	17.501	ug/l	96
31) Cyclohexane	5.471	56	3619	2.026	ug/l #	96
39) Methylcyclohexane	7.385	83	4568	2.459	ug/l	89
40) Benzene	6.037	78	730547	150.978	ug/l	100
43) Isopropyl Acetate	6.336	43	7217	2.859	ug/l #	92
52) Toluene	8.720	92	415384	132.774	ug/l	100
67) Ethyl Benzene	10.195	91	28727	4.589	ug/l	97
68) m/p-Xylenes	10.299	106	71020	29.162	ug/l	98
69) o-Xylene	10.640	106	21468	8.918	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	11332	2.178	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	19483	3.751	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
Data File : VX033508.D
Acq On : 20 Dec 2022 15:17
Operator : JC/MD
Sample : N6149-06 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WANA-1

Quant Time: Dec 21 04:15:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

