

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
 Data File : VX033520.D
 Acq On : 20 Dec 2022 19:53
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 21 04:17:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136465	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233601	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206062	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	38002	43.407	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.820%
35) Dibromofluoromethane	5.391	113	45934	43.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.440%
50) Toluene-d8	8.647	98	101078	45.987	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.980%#
62) 4-Bromofluorobenzene	11.079	95	81482	48.483	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
Data File : VX033520.D
Acq On : 20 Dec 2022 19:53
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 21 04:17:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

