


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122022\
Data File : VX033497.D
Acq On : 20 Dec 2022 10:34
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Title : SW846 8260
Last Update : Wed Dec 07 12:43:54 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	6446	PASS
75	95	30	60	53.1	18274	PASS
95	95	100	100	100.0	34432	PASS
96	95	5	9	6.7	2298	PASS
173	174	0.00	2	1.1	315	PASS
174	95	50	100	80.7	27787	PASS
175	174	5	9	7.2	2003	PASS
176	174	95	101	96.5	26821	PASS
177	176	5	9	6.8	1831	PASS