

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122023\
 Data File : VX039394.D
 Acq On : 20 Dec 2023 12:25
 Operator : JC/MD
 Sample : 05869-08 100X
 Misc : 3.72g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AT4

Quant Time: Dec 21 00:25:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Dec 20 00:14:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	168951	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	156197	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	72100	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	38834	29.305	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	58.620%#
7) Chloroethane-d5	1.666	69	36329	37.090	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.180%
11) 1,1-Dichloroethene-d2	2.306	65	19090	32.818	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.640%
21) 2-Butanone-d5	4.446	46	117117	97.636	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.640%
24) Chloroform-d	5.056	84	100151	42.693	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.380%
26) 1,2-Dichloroethane-d4	5.952	65	79472	52.310	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	104.620%
32) Benzene-d6	5.976	84	197126	40.893	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.780%
36) 1,2-Dichloropropane-d6	7.305	67	70123	45.201	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.400%
41) Toluene-d8	8.646	98	158772	36.344	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.680%#
43) trans-1,3-Dichloroprop...	8.951	79	28843	36.535	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.080%
47) 2-Hexanone-d5	9.384	63	83933	90.584	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.580%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	103054	46.827	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.660%
66) 1,2-Dichlorobenzene-d4	12.317	152	54348	39.879	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.760%#
Target Compounds						
					Qvalue	
51) Chlorobenzene	10.079	112	2765	0.850	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122023\
Data File : VX039394.D
Acq On : 20 Dec 2023 12:25
Operator : JC/MD
Sample : 05869-08 100X
Misc : 3.72g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AT4

Quant Time: Dec 21 00:25:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Dec 20 00:14:00 2023
Response via : Initial Calibration

