

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044444.D
 Acq On : 20 Dec 2024 12:03
 Operator : JC/MD
 Sample : P5281-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D1-20241212DL

Quant Time: Dec 20 22:24:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	118538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211612	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184574	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	89218	42.927	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.860%
35) Dibromofluoromethane	5.385	113	70747	48.272	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.540%
50) Toluene-d8	8.647	98	250889	50.741	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.480%
62) 4-Bromofluorobenzene	11.079	95	79951	46.275	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.540%
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	3329	2.301	ug/l	98
12) 1,1-Dichloroethene	2.313	96	505	0.380	ug/l #	88
27) cis-1,2-Dichloroethene	4.501	96	609	0.340	ug/l	94
44) Trichloroethene	7.123	130	58695	40.920	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044444.D
Acq On : 20 Dec 2024 12:03
Operator : JC/MD
Sample : P5281-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D1-20241212DL

Quant Time: Dec 20 22:24:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

