

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044445.D
Acq On : 20 Dec 2024 12:26
Operator : JC/MD
Sample : P5314-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-TB-20241213

Quant Time: Dec 20 14:08:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117644	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	208582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	178072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	71905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88659	42.982	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.960%
35) Dibromofluoromethane	5.385	113	66732	46.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.647	98	246144	50.504	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.000%
62) 4-Bromofluorobenzene	11.079	95	80161	47.070	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.140%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044445.D
Acq On : 20 Dec 2024 12:26
Operator : JC/MD
Sample : P5314-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-TB-20241213

Quant Time: Dec 20 14:08:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

