

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044446.D
 Acq On : 20 Dec 2024 12:49
 Operator : JC/MD
 Sample : P5314-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190A-GW-698-700

Quant Time: Dec 20 14:08:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211457	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	69187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90187	42.254	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.500%
35) Dibromofluoromethane	5.385	113	68986	47.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	252545	51.113	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.220%
62) 4-Bromofluorobenzene	11.079	95	79964	46.316	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.640%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.940	59	9223	30.660	ug/l #	72
16) Acetone	2.386	43	9498	10.915	ug/l	98
25) 2-Butanone	4.617	43	2159	1.745	ug/l	99
44) Trichloroethene	7.122	130	6238	4.352	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044446.D
Acq On : 20 Dec 2024 12:49
Operator : JC/MD
Sample : P5314-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-GW-698-700

Quant Time: Dec 20 14:08:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

