

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044447.D
 Acq On : 20 Dec 2024 13:13
 Operator : JC/MD
 Sample : P5314-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-190A-GW-678-680

Quant Time: Dec 20 14:08:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :John Carlone 12/20/2024
 Supervised By :Semsettin Yesilyurt 12/23/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	116922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	205157	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179312	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	74123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87608	42.735	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	85.480%		
35) Dibromofluoromethane	5.379	113	68550	48.244	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.480%		
50) Toluene-d8	8.647	98	246944	51.514	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	103.020%		
62) 4-Bromofluorobenzene	11.079	95	83804	50.031	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.060%		
Target Compounds						
9) 1,1,2-Trichlorotrifluo...	2.319	101	7662	5.370	ug/l	96
12) 1,1-Dichloroethene	2.312	96	1852	1.413	ug/l #	77
16) Acetone	2.386	43	4141	4.955	ug/l	91
27) cis-1,2-Dichloroethene	4.501	96	2816m	1.592	ug/l	
30) Chloroform	5.092	83	2628	0.854	ug/l #	79
38) Carbon Tetrachloride	5.684	117	2826	1.429	ug/l #	73
44) Trichloroethene	7.123	130	401769	288.909	ug/l	97
55) 1,1,2-Trichloroethane	9.159	97	1835	1.319	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044447.D
Acq On : 20 Dec 2024 13:13
Operator : JC/MD
Sample : P5314-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-190A-GW-678-680

Quant Time: Dec 20 14:08:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 12/20/2024
Supervised By :Semsettin Yesilyurt 12/23/2024

