

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044448.D
 Acq On : 20 Dec 2024 13:36
 Operator : JC/MD
 Sample : P5314-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB190A-HYD-20241217

Quant Time: Dec 20 14:08:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122994	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	213253	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	188396	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	78770	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90570	41.999	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.000%		
35) Dibromofluoromethane	5.385	113	70242	47.558	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.120%		
50) Toluene-d8	8.647	98	255708	51.317	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.640%		
62) 4-Bromofluorobenzene	11.079	95	85988	49.386	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	98.780%		
Target Compounds						
16) Acetone	2.386	43	2376	2.703	ug/l #	82
44) Trichloroethene	7.135	130	2039	1.411	ug/l	85
47) Bromodichloromethane	7.824	83	449	0.226	ug/l #	76
60) Dibromochloromethane	9.525	129	368	1.476	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044448.D
Acq On : 20 Dec 2024 13:36
Operator : JC/MD
Sample : P5314-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB190A-HYD-20241217

Quant Time: Dec 20 14:08:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

