

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044451.D
 Acq On : 20 Dec 2024 14:46
 Operator : JC/MD
 Sample : P5281-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW1-7-20241212

Quant Time: Dec 20 22:25:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	117082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	205938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173219	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	86944	42.353	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.700%
35) Dibromofluoromethane	5.379	113	67575	47.378	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.760%
50) Toluene-d8	8.647	98	242944	50.488	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.980%
62) 4-Bromofluorobenzene	11.079	95	73620	43.785	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	87.560%
Target Compounds						
44) Trichloroethene	7.147	130	849	0.608	ug/l	Qvalue 36

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044451.D
Acq On : 20 Dec 2024 14:46
Operator : JC/MD
Sample : P5281-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW1-7-20241212

Quant Time: Dec 20 22:25:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

