

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044454.D
 Acq On : 20 Dec 2024 15:56
 Operator : JC/MD
 Sample : P5314-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB190A-HYD-20241217

Quant Time: Dec 20 16:22:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	120419	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	211996	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	183684	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90307	42.772	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.540%
35) Dibromofluoromethane	5.385	113	68667	46.767	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.540%
50) Toluene-d8	8.647	98	250790	50.629	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.260%
62) 4-Bromofluorobenzene	11.079	95	78660	45.445	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	90.900%
Target Compounds						
16) Acetone	2.386	43	2305	2.678	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044454.D
Acq On : 20 Dec 2024 15:56
Operator : JC/MD
Sample : P5314-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB190A-HYD-20241217

Quant Time: Dec 20 16:22:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

