

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044455.D
 Acq On : 20 Dec 2024 16:19
 Operator : JC/MD
 Sample : P5281-06DL 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20241213DL

Quant Time: Dec 20 22:27:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121258	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216230	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191670	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	80370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	90673	42.649	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.300%
35) Dibromofluoromethane	5.379	113	69035	46.097	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.200%
50) Toluene-d8	8.647	98	254320	50.336	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.680%
62) 4-Bromofluorobenzene	11.079	95	80586	45.646	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.300%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	2415	1.632	ug/l	98
12) 1,1-Dichloroethene	2.313	96	809	0.595	ug/l	96
16) Acetone	2.386	43	348	0.401	ug/l	97
44) Trichloroethene	7.123	130	49142	33.528	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044455.D
Acq On : 20 Dec 2024 16:19
Operator : JC/MD
Sample : P5281-06DL 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20241213DL

Quant Time: Dec 20 22:27:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

