

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044457.D
 Acq On : 20 Dec 2024 17:06
 Operator : JC/MD
 Sample : PB165734TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB165734TB

Quant Time: Dec 20 22:28:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	206131	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	72950	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85857	42.309	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	84.620%		
35) Dibromofluoromethane	5.385	113	68907	48.266	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.540%		
50) Toluene-d8	8.647	98	247233	51.331	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.660%		
62) 4-Bromofluorobenzene	11.079	95	77726	46.183	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.360%		
Target Compounds						
16) Acetone	2.386	43	29329	35.451	ug/l	93
18) Methyl Acetate	2.703	43	3429	1.608	ug/l #	89
20) Methylene Chloride	2.788	84	3206	2.063	ug/l #	78
30) Chloroform	5.086	83	453535	148.828	ug/l	99
43) Isopropyl Acetate	6.342	43	108014	30.295	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044457.D
Acq On : 20 Dec 2024 17:06
Operator : JC/MD
Sample : PB165734TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB165734TB

Quant Time: Dec 20 22:28:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

