

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044458.D
 Acq On : 20 Dec 2024 17:30
 Operator : JC/MD
 Sample : P5280-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-11-20241212

Quant Time: Dec 20 22:29:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	195214	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65426	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85210	42.611	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	85.220%		
35) Dibromofluoromethane	5.385	113	65619	48.534	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.060%		
50) Toluene-d8	8.647	98	233939	51.287	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	102.580%		
62) 4-Bromofluorobenzene	11.079	95	70521	44.246	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	88.500%		
Target Compounds						
16) Acetone	2.386	43	1286	1.577	ug/l #	77
18) Methyl Acetate	2.709	43	4129	1.964	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044458.D
Acq On : 20 Dec 2024 17:30
Operator : JC/MD
Sample : P5280-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BPOW6-11-20241212

Quant Time: Dec 20 22:29:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

