

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
 Data File : VX044463.D
 Acq On : 20 Dec 2024 19:26
 Operator : JC/MD
 Sample : P5355-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR251688

Quant Time: Dec 20 22:36:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 12 04:28:14 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	124489	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	216029	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	184635	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88705	40.640	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	81.280%	
35) Dibromofluoromethane	5.385	113	70904	47.389	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.780%	
50) Toluene-d8	8.646	98	257204	50.954	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.900%	
62) 4-Bromofluorobenzene	11.079	95	79235	44.923	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	89.840%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	28565	32.101	ug/l	99
18) Methyl Acetate	2.709	43	2479	1.080	ug/l	99
20) Methylene Chloride	2.782	84	3227	1.931	ug/l	85
30) Chloroform	5.086	83	21755	6.637	ug/l	93
43) Isopropyl Acetate	6.342	43	77071	20.626	ug/l	99
44) Trichloroethene	7.122	130	2438	1.665	ug/l	99
95) Naphthalene	13.780	128	6838	1.222	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122024\
Data File : VX044463.D
Acq On : 20 Dec 2024 19:26
Operator : JC/MD
Sample : P5355-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR251688

Quant Time: Dec 20 22:36:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121124W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 12 04:28:14 2024
Response via : Initial Calibration

