

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006724.D
 Acq On : 21 Dec 2018 15:29
 Operator : JC/MD
 Sample : J6514-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0495DL

Quant Time: Dec 22 04:38:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	644470	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	999686	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	897765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	424024	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	341714	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
35) Dibromofluoromethane	5.51	113	302455	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
50) Toluene-d8	8.71	98	1159644	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
62) 4-Bromofluorobenzene	11.14	95	413410	47.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.72%	
Target Compounds						
4) Vinyl Chloride	1.41	62	10686	1.611	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	59484	8.450	ug/l	91
44) Trichloroethene	7.21	130	207402	27.227	ug/l	98
64) Tetrachloroethene	9.34	164	2016	0.284	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006724.D
Acq On : 21 Dec 2018 15:29
Operator : JC/MD
Sample : J6514-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
S49-0495DL

Quant Time: Dec 22 04:38:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

