

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006743.D
 Acq On : 22 Dec 2018 00:51
 Operator : JC/MD
 Sample : J6191-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NB-07-122018-B

Quant Time: Dec 22 05:45:54 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	627806	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	991500	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	905069	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	426459	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	345511	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
35) Dibromofluoromethane	5.51	113	304105	48.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.00%	
50) Toluene-d8	8.71	98	1164603	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
62) 4-Bromofluorobenzene	11.14	95	406211	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
Target Compounds						
16) Acetone	2.45	43	47970	15.431	ug/l	97
18) Methyl Acetate	2.78	43	23961	2.621	ug/l #	90
20) Methylene Chloride	2.86	84	2410004	374.919	ug/l	96
43) Isopropyl Acetate	6.46	43	29332	2.255	ug/l	99
95) Naphthalene	13.83	128	347846	11.111	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006743.D
Acq On : 22 Dec 2018 00:51
Operator : JC/MD
Sample : J6191-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NB-07-122018-B

Quant Time: Dec 22 05:45:54 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

