

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006744.D
 Acq On : 22 Dec 2018 01:18
 Operator : JC/MD
 Sample : J6192-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HD-01-122018

Quant Time: Dec 22 05:47:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	636289	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	988409	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	893785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	415872	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	341091	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
35) Dibromofluoromethane	5.51	113	297313	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
50) Toluene-d8	8.71	98	1159464	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.14	95	402217	46.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.20%	
Target Compounds						
16) Acetone	2.45	43	84407	26.791	ug/l	99
18) Methyl Acetate	2.78	43	24729	2.669	ug/l	98
20) Methylene Chloride	2.86	84	298785	45.862	ug/l	99
43) Isopropyl Acetate	6.46	43	30416	2.346	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006744.D
Acq On : 22 Dec 2018 01:18
Operator : JC/MD
Sample : J6192-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HD-01-122018

Quant Time: Dec 22 05:47:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

