

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006747.D
 Acq On : 22 Dec 2018 02:38
 Operator : JC/MD
 Sample : J6200-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-05-122018

Quant Time: Dec 22 05:53:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	632742	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1000485	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	891514	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	415449	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	341723	50.11	ug/l	0.00
Spiked Amount			Recovery	=	100.22%	
35) Dibromofluoromethane	5.51	113	300521	47.50	ug/l	0.00
Spiked Amount			Recovery	=	95.00%	
50) Toluene-d8	8.71	98	1169169	49.02	ug/l	0.00
Spiked Amount			Recovery	=	98.04%	
62) 4-Bromofluorobenzene	11.13	95	393362	45.03	ug/l	0.00
Spiked Amount			Recovery	=	90.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	53025	16.924	ug/l	97
18) Methyl Acetate	2.78	43	14843	1.611	ug/l	98
20) Methylene Chloride	2.86	84	48108	7.426	ug/l	94
43) Isopropyl Acetate	6.46	43	30875	2.353	ug/l	95
52) Toluene	8.78	92	239287	14.705	ug/l	98
67) Ethyl Benzene	10.25	91	137936	4.476	ug/l	97
68) m/p-Xylenes	10.36	106	229665	18.885	ug/l	99
69) o-Xylene	10.70	106	95166	7.924	ug/l	100
78) n-propylbenzene	11.36	91	54608	1.745	ug/l	97
80) 1,3,5-Trimethylbenzene	11.51	105	102474	4.377	ug/l	97
84) 1,2,4-Trimethylbenzene	11.80	105	362938	14.740	ug/l	100
95) Naphthalene	13.83	128	117423	3.850	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006747.D
Acq On : 22 Dec 2018 02:38
Operator : JC/MD
Sample : J6200-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-05-122018

Quant Time: Dec 22 05:53:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

