

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006749.D
 Acq On : 22 Dec 2018 03:32
 Operator : JC/MD
 Sample : J6520-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ROLLOFF-165

Manual Integrations
 APPROVED

MMDadoda
 12/22/2018 11:20:21 AM

Quant Time: Dec 22 05:57:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	660300	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1022458	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	938951	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	429989	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	353258	49.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.28%	
35) Dibromofluoromethane	5.51	113	315475	48.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.58%	
50) Toluene-d8	8.71	98	1213259	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
62) 4-Bromofluorobenzene	11.13	95	420401	47.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.18%	
Target Compounds						
16) Acetone	2.45	43	45069	13.785	ug/l	92
20) Methylene Chloride	2.83	84	39032424m	5773.364	ug/l	
43) Isopropyl Acetate	6.46	43	30272	2.257	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006749.D
Acq On : 22 Dec 2018 03:32
Operator : JC/MD
Sample : J6520-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ROLLOFF-165

Manual Integrations
APPROVED

MMDadoda
12/22/2018 11:20:21 AM

Quant Time: Dec 22 05:57:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

