

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006755.D
 Acq On : 22 Dec 2018 06:12
 Operator : JC/MD
 Sample : J6503-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-GROUNDWATER-1-1

Quant Time: Dec 22 09:42:22 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	631014	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	986150	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	882807	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	416668	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	344171	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
35) Dibromofluoromethane	5.51	113	299396	48.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.02%	
50) Toluene-d8	8.71	98	1152401	49.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.04%	
62) 4-Bromofluorobenzene	11.14	95	403248	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
Target Compounds						
16) Acetone	2.45	43	51617	16.520	ug/l	98
17) Carbon Disulfide	2.57	76	12801	0.858	ug/l #	95
20) Methylene Chloride	2.86	84	16601	2.569	ug/l	90
25) 2-Butanone	4.70	43	16513	3.757	ug/l	94
52) Toluene	8.79	92	157066	9.793	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006755.D
Acq On : 22 Dec 2018 06:12
Operator : JC/MD
Sample : J6503-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-GROUNDWATER-1-1

Quant Time: Dec 22 09:42:22 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

