

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122118\
 Data File : VX006757.D
 Acq On : 22 Dec 2018 07:05
 Operator : JC/MD
 Sample : J6496-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WATER-1

Quant Time: Dec 22 09:50:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	636659	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	997829	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	893599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	422706	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	350731	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
35) Dibromofluoromethane	5.51	113	302029	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
50) Toluene-d8	8.71	98	1169764	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
62) 4-Bromofluorobenzene	11.14	95	411120	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	1322679	419.572	ug/l	100
18) Methyl Acetate	2.78	43	61563	6.641	ug/l	94
20) Methylene Chloride	2.87	84	12526	1.922	ug/l	94
25) 2-Butanone	4.70	43	567822	128.052	ug/l	99
30) Chloroform	5.22	83	6169	0.570	ug/l	100
51) 4-Methyl-2-Pentanone	8.65	43	40809	4.901	ug/l	97
59) 2-Hexanone	9.49	43	78163	12.268	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122118\
Data File : VX006757.D
Acq On : 22 Dec 2018 07:05
Operator : JC/MD
Sample : J6496-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WATER-1

Quant Time: Dec 22 09:50:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

