

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122120\
 Data File : VX020257.D
 Acq On : 21 Dec 2020 17:59
 Operator : JC/MD
 Sample : L5123-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-20201211

Quant Time: Dec 22 01:35:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	244827	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.830	114	416278	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.098	117	391582	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.061	152	165304	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.031	65	168599	51.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.74%	
35) Dibromofluoromethane	5.464	113	134345	50.45	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.90%	
50) Toluene-d8	8.696	98	504888	49.26	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 98.52%	
62) 4-Bromofluorobenzene	11.122	95	177052	45.13	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 90.26%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122120\
Data File : VX020257.D
Acq On : 21 Dec 2020 17:59
Operator : JC/MD
Sample : L5123-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-20201211

Quant Time: Dec 22 01:35:37 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

