

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122122\
 Data File : VX033529.D
 Acq On : 21 Dec 2022 12:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Dec 22 03:57:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133377	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182355	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	62786	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	37272	43.559	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.120%
35) Dibromofluoromethane	5.385	113	47570	50.358	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.720%
50) Toluene-d8	8.647	98	93558	47.344	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.680%
62) 4-Bromofluorobenzene	11.079	95	66243	43.841	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122122\
Data File : VX033529.D
Acq On : 21 Dec 2022 12:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Dec 22 03:57:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

