

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122122\
 Data File : VX033530.D
 Acq On : 21 Dec 2022 13:11
 Operator : JC/MD
 Sample : PB149803TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB149803TB

Quant Time: Dec 22 03:57:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 07 12:43:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	135858	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	228665	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100374	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	40914	46.943	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.880%
35) Dibromofluoromethane	5.385	113	47961	46.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.260%
50) Toluene-d8	8.647	98	97771	45.442	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.880%#
62) 4-Bromofluorobenzene	11.079	95	81127	49.314	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	15170	23.521	ug/l	98
20) Methylene Chloride	2.788	84	7348	5.109	ug/l	94
37) Ethyl Acetate	4.715	43	16372	8.637	ug/l	97
43) Isopropyl Acetate	6.342	43	74929	24.568	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122122\
Data File : VX033530.D
Acq On : 21 Dec 2022 13:11
Operator : JC/MD
Sample : PB149803TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB149803TB

Quant Time: Dec 22 03:57:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 07 12:43:54 2022
Response via : Initial Calibration

