



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122122\
Data File : VX033522.D
Acq On : 21 Dec 2022 09:40
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X120722W.M
Title : SW846 8260
Last Update : Wed Dec 07 12:43:54 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.1	8278	PASS
75	95	30	60	48.8	25059	PASS
95	95	100	100	100.0	51352	PASS
96	95	5	9	6.6	3366	PASS
173	174	0.00	2	0.9	368	PASS
174	95	50	100	80.6	41403	PASS
175	174	5	9	7.4	3069	PASS
176	174	95	101	95.7	39608	PASS
177	176	5	9	6.9	2742	PASS