

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122123\
 Data File : VX039416.D
 Acq On : 21 Dec 2023 09:53
 Operator : JC/MD
 Sample : VX1221MBL01
 Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK656

Quant Time: Dec 22 00:41:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 22 00:41:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	192740	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.049	117	176323	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	87502	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	68241	45.140	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.280%
7) Chloroethane-d5	1.672	69	51551	46.135	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	92.260%
11) 1,1-Dichloroethene-d2	2.306	65	33355	50.264	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.520%
21) 2-Butanone-d5	4.440	46	135604	99.095	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.090%
24) Chloroform-d	5.050	84	131254	49.046	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.100%
26) 1,2-Dichloroethane-d4	5.946	65	93718	54.074	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.140%
32) Benzene-d6	5.970	84	266842	49.037	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.080%
36) 1,2-Dichloropropane-d6	7.306	67	87954	50.224	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.440%
41) Toluene-d8	8.647	98	236506	47.959	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.920%
43) trans-1,3-Dichloroprop...	8.946	79	39630	44.469	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.940%
47) 2-Hexanone-d5	9.378	63	96461	92.222	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	92.220%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	125710	50.602	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.200%
66) 1,2-Dichlorobenzene-d4	12.317	152	83317	50.375	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.740%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122123\
Data File : VX039416.D
Acq On : 21 Dec 2023 09:53
Operator : JC/MD
Sample : VX1221MBL01
Misc : 5.00g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK656

Quant Time: Dec 22 00:41:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM120523WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 22 00:41:09 2023
Response via : Initial Calibration

