

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122123\
 Data File : VX039426.D
 Acq On : 21 Dec 2023 13:56
 Operator : JC/MD
 Sample : 05835-10 4X
 Misc : 3.22g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AQ7

Quant Time: Dec 22 00:44:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Dec 22 00:41:09 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	184554	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	174459	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	86846	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	57361	39.627	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	79.260%
7) Chloroethane-d5	1.648	69	38160	35.665	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	71.340%
11) 1,1-Dichloroethene-d2	2.300	65	29395	46.261	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	92.520%
21) 2-Butanone-d5	4.452	46	138266	105.522	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	105.520%
24) Chloroform-d	5.050	84	124065	48.416	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.840%
26) 1,2-Dichloroethane-d4	5.952	65	91002	54.836	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.680%
32) Benzene-d6	5.970	84	255125	47.384	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.760%
36) 1,2-Dichloropropane-d6	7.306	67	85600	49.402	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.800%
41) Toluene-d8	8.647	98	223569	45.820	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.640%
43) trans-1,3-Dichloroprop...	8.952	79	35817	40.620	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.240%
47) 2-Hexanone-d5	9.384	63	104058	100.549	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	100.550%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	125264	50.961	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.920%
66) 1,2-Dichlorobenzene-d4	12.323	152	84589	51.530	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.060%
Target Compounds						
					Qvalue	
51) Chlorobenzene	10.079	112	295371	81.335	ug/L	98
65) 1,4-Dichlorobenzene	12.043	146	6638	2.412	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122123\
Data File : VX039426.D
Acq On : 21 Dec 2023 13:56
Operator : JC/MD
Sample : 05835-10 4X
Misc : 3.22g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AQ7

Quant Time: Dec 22 00:44:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML120523WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Dec 22 00:41:09 2023
Response via : Initial Calibration

