

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122218\
 Data File : VX006776.D
 Acq On : 22 Dec 2018 17:03
 Operator : JC/MD
 Sample : J6514-04DL2 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0499DL2

Quant Time: Dec 24 05:46:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	623403	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	977727	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	884587	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	413466	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	338311	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
35) Dibromofluoromethane	5.50	113	290438	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
50) Toluene-d8	8.71	98	1146943	49.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.40%	
62) 4-Bromofluorobenzene	11.13	95	404262	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
Target Compounds						
4) Vinyl Chloride	1.41	62	17460	2.721	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	5994	0.971	ug/l	85
27) cis-1,2-Dichloroethene	4.60	96	76995	11.308	ug/l	91
44) Trichloroethene	7.21	130	304734	40.903	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122218\
Data File : VX006776.D
Acq On : 22 Dec 2018 17:03
Operator : JC/MD
Sample : J6514-04DL2 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
S49-0499DL2

Quant Time: Dec 24 05:46:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration

