

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX122218\
 Data File : VX006778.D
 Acq On : 22 Dec 2018 17:56
 Operator : JC/MD
 Sample : J6497-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Dec 24 12:03:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 19 15:20:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	613135	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	960238	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	883129	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	417144	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	334995	50.69	ug/l	0.00
Spiked Amount			Recovery	=	101.38%	
35) Dibromofluoromethane	5.51	113	289692	47.71	ug/l	0.00
Spiked Amount			Recovery	=	95.42%	
50) Toluene-d8	8.71	98	1137217	49.68	ug/l	0.00
Spiked Amount			Recovery	=	99.36%	
62) 4-Bromofluorobenzene	11.13	95	404948	48.30	ug/l	0.00
Spiked Amount			Recovery	=	96.60%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX122218\
Data File : VX006778.D
Acq On : 22 Dec 2018 17:56
Operator : JC/MD
Sample : J6497-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Dec 24 12:03:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X121818W.M
Quant Title : SW846 8260
QLast Update : Wed Dec 19 15:20:58 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX006778.D

Acq: 22 Dec 2018 17:56

Instrument :

MSVOA_X

ClientSampleId :


MW-3

Tot Ion:168 Resp: 613135

Ion Ratio Lower Upper

168 100

99 52.6 41.8 62.8



Abundance Ion 167.90 (167.60 to 168.60): VX006591.D (-725) (-)

Ion 98.90 (98.60 to 99.60): VX006591.D (-725) (-)

5.67

200000



150000

100000

50000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 50.695 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.01 min

Lab File: VX006778.D


Acq: 22 Dec 2018 17:56

Tgt Ion: 65 Resp: 334995

Ion Ratio Lower Upper

65 100

67 52.8 0.0 105.0



Abundance Ion 64.90 (64.60 to 65.60): VX006591.D (-790) (-)

Ion 66.90 (66.60 to 67.60): VX006591.D (-790) (-)

6.07


150000


100000

50000

0

Time-->





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.00 min

Lab File: VX006778.D

Acq: 22 Dec 2018 17:56

Instrument :

MSVOA_X

ClientSampled :

MW-3


Tot Ion:114 Resp: 960238

Ion Ratio Lower Upper

114 100

63 18.6 0.0 34.8




88 14.8 0.0 28.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.707 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX006778.D

Acq: 22 Dec 2018 17:56


Tgt Ion:113 Resp: 289692

Ion Ratio Lower Upper

113 100

111 101.5 81.6 122.4



192 20.3 16.7 25.1




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.675 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. -0.00 min

Lab File: VX006778.D

Acq: 22 Dec 2018 17:56

Instrument :

MSVOA_X

ClientSampleId :





MW-3

Tot Ion: 98 Resp: 1137217

Ion Ratio Lower Upper

98 100

100 64.6 52.0 78.0



#62

4-Bromofluorobenzene

Concen: 48.298 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX006778.D

Acq: 22 Dec 2018 17:56




Tgt Ion: 95 Resp: 404948


Ion Ratio Lower Upper

95 100

174 87.5 0.0 182.2

176 85.3 0.0 178.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX006778.D
Acq: 22 Dec 2018 17:56

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Tot Ion:117 Resp: 883129
Ion Ratio Lower Upper
117 100
82 52.3 41.0 61.6
119 31.2 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX006778.D
Acq: 22 Dec 2018 17:56

Tgt Ion:152 Resp: 417144
Ion Ratio Lower Upper
152 100
115 55.9 39.1 117.2
150 156.7 0.0 344.8

