

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX122220\
 Data File : VX020290.D
 Acq On : 22 Dec 2020 08:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 00:25:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	144	0.00
2 T	Dichlorodifluoromethane	50.000	48.654	2.7	149	0.00
3 P	Chloromethane	50.000	42.002	16.0	123	0.00
4 C	Vinyl Chloride	50.000	44.480	11.0#	137	0.00
5 T	Bromomethane	50.000	42.028	15.9	127	0.00
6 T	Chloroethane	50.000	44.628	10.7	128	0.00
7 T	Trichlorofluoromethane	50.000	47.142	5.7	148	0.00
8 T	Diethyl Ether	50.000	44.048	11.9	118	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	46.414	7.2	146	0.00
10 T	Methyl Iodide	50.000	38.624	22.8	111	0.00
11 T	Tert butyl alcohol	250.000	224.462	10.2	134	0.00
12 CM	1,1-Dichloroethene	50.000	45.037	9.9#	139	0.00
13 T	Acrolein	250.000	196.595	21.4	112	0.00
14 T	Allyl chloride	50.000	45.470	9.1	129	0.00
15 T	Acrylonitrile	250.000	224.588	10.2	121	0.00
16 T	Acetone	250.000	289.156	-15.7	165	0.00
17 T	Carbon Disulfide	50.000	47.217	5.6	146	0.00
18 T	Methyl Acetate	50.000	46.332	7.3	126	0.00
19 T	Methyl tert-butyl Ether	50.000	46.483	7.0	122	0.00
20 T	Methylene Chloride	50.000	43.865	12.3	124	0.00
21 T	trans-1,2-Dichloroethene	50.000	44.677	10.6	130	0.00
22 T	Diisopropyl ether	50.000	45.671	8.7	121	0.00
23 T	Vinyl Acetate	250.000	234.624	6.2	120	0.00
24 P	1,1-Dichloroethane	50.000	45.709	8.6	128	0.00
25 T	2-Butanone	250.000	260.756	-4.3	141	0.00
26 T	2,2-Dichloropropane	50.000	47.831	4.3	139	0.00
27 T	cis-1,2-Dichloroethene	50.000	45.161	9.7	125	0.00
28 T	Bromochloromethane	50.000	46.046	7.9	128	0.00
29 T	Tetrahydrofuran	250.000	225.634	9.7	120	0.00
30 C	Chloroform	50.000	45.436	9.1#	125	0.00
31 T	Cyclohexane	50.000	47.071	5.9	146	0.00
32 T	1,1,1-Trichloroethane	50.000	47.202	5.6	139	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.379	7.2	128	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	133	0.00
35 S	Dibromofluoromethane	50.000	49.250	1.5	131	0.00
36 T	1,1-Dichloropropene	50.000	47.647	4.7	139	0.00
37 T	Ethyl Acetate	50.000	47.001	6.0	122	0.00
38 T	Carbon Tetrachloride	50.000	49.914	0.2	143	0.00
39 T	Methylcyclohexane	50.000	51.374	-2.7	156	0.00
40 TM	Benzene	50.000	47.879	4.2	128	0.00
41 T	Methacrylonitrile	50.000	48.007	4.0	119	0.00
42 TM	1,2-Dichloroethane	50.000	46.807	6.4	121	0.00
43 T	Isopropyl Acetate	50.000	48.008	4.0	122	0.00
44 TM	Trichloroethene	50.000	46.819	6.4	132	0.00
45 C	1,2-Dichloropropane	50.000	47.249	5.5#	123	0.00
46 T	Dibromomethane	50.000	45.775	8.5	119	0.00
47 T	Bromodichloromethane	50.000	49.452	1.1	126	0.00
48 T	Methyl methacrylate	50.000	49.195	1.6	123	0.00
49 T	1,4-Dioxane	1000.000	983.345	1.7	126	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122220\
 Data File : VX020290.D
 Acq On : 22 Dec 2020 08:31
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 00:25:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : Sw846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50 S	Toluene-d8	50.000	48.822	2.4	132	0.00
51 T	4-Methyl-2-Pentanone	250.000	246.578	1.4	124	0.00
52 CM	Toluene	50.000	49.523	1.0#	131	0.00
53 T	t-1,3-Dichloropropene	50.000	50.698	-1.4	127	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.643	0.7	124	0.00
55 T	1,1,2-Trichloroethane	50.000	47.338	5.3	122	0.00
56 T	Ethyl methacrylate	50.000	51.218	-2.4	125	0.00
57 T	1,3-Dichloropropane	50.000	47.012	6.0	120	0.00
58 T	2-Chloroethyl vinyl ether	250.000	224.822	10.1	114	0.00
59 T	2-Hexanone	250.000	278.330	-11.3	140	0.00
60 T	Dibromochloromethane	50.000	50.540	-1.1	127	0.00
61 T	1,2-Dibromoethane	50.000	47.878	4.2	122	0.00
62 S	4-Bromofluorobenzene	50.000	50.057	-0.1	135	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	128	0.00
64 T	Tetrachloroethene	50.000	46.645	6.7	134	0.00
65 PM	Chlorobenzene	50.000	47.874	4.3	126	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.913	0.2	128	0.00
67 C	Ethyl Benzene	50.000	49.701	0.6#	131	0.00
68 T	m/p-Xylenes	100.000	100.016	-0.0	129	0.00
69 T	o-Xylene	50.000	49.260	1.5	125	0.00
70 T	Styrene	50.000	50.705	-1.4	125	0.00
71 P	Bromoform	50.000	52.757	-5.5	135	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	131	0.00
73 T	Isopropylbenzene	50.000	49.421	1.2	135	0.00
74 T	N-aryl acetate	50.000	51.278	-2.6	128	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	45.979	8.0	122	0.00
76 T	1,2,3-Trichloropropane	50.000	46.897	6.2	125	0.00
77 T	Bromobenzene	50.000	46.182	7.6	127	0.00
78 T	n-propylbenzene	50.000	49.976	0.0	137	0.00
79 T	2-Chlorotoluene	50.000	47.847	4.3	129	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.986	0.0	133	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.795	-1.6	129	0.00
82 T	4-Chlorotoluene	50.000	48.693	2.6	130	0.00
83 T	tert-Butylbenzene	50.000	49.492	1.0	136	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.681	0.6	129	0.00
85 T	sec-Butylbenzene	50.000	51.671	-3.3	146	0.00
86 T	p-Isopropyltoluene	50.000	51.085	-2.2	139	0.00
87 T	1,3-Dichlorobenzene	50.000	46.123	7.8	124	0.00
88 T	1,4-Dichlorobenzene	50.000	44.688	10.6	124	0.00
89 T	n-Butylbenzene	50.000	53.213	-6.4	148	0.00
90 T	Hexachloroethane	50.000	51.381	-2.8	145	0.00
91 T	1,2-Dichlorobenzene	50.000	44.969	10.1	122	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.663	4.7	122	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.273	3.5	131	0.00
94 T	Hexachlorobutadiene	50.000	51.804	-3.6	165	0.00
95 T	Naphthalene	50.000	49.376	1.2	124	0.00
96 T	1,2,3-Trichlorobenzene	50.000	46.580	6.8	127	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122220\
Data File : VX020290.D
Acq On : 22 Dec 2020 08:31
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 23 00:25:54 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6