

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX122220\
 Data File : VX020316.D
 Acq On : 22 Dec 2020 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 00:36:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	112	0.00
2 T	Dichlorodifluoromethane	50.000	49.819	0.4	119	0.00
3 P	Chloromethane	50.000	47.937	4.1	109	0.00
4 C	Vinyl Chloride	50.000	48.772	2.5#	117	0.00
5 T	Bromomethane	50.000	49.068	1.9	116	0.00
6 T	Chloroethane	50.000	49.818	0.4	111	0.00
7 T	Trichlorofluoromethane	50.000	50.744	-1.5	124	0.00
8 T	Diethyl Ether	50.000	49.019	2.0	103	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.000	-2.0	125	0.00
10 T	Methyl Iodide	50.000	49.181	1.6	112	0.00
11 T	Tert butyl alcohol	250.000	257.302	-2.9	120	0.00
12 CM	1,1-Dichloroethene	50.000	51.918	-3.8#	125	0.00
13 T	Acrolein	250.000	351.514	-40.6#	156	0.00
14 T	Allyl chloride	50.000	49.569	0.9	110	0.00
15 T	Acrylonitrile	250.000	263.346	-5.3	111	0.00
16 T	Acetone	250.000	232.491	7.0	104	0.00
17 T	Carbon Disulfide	50.000	48.895	2.2	118	0.00
18 T	Methyl Acetate	50.000	54.213	-8.4	115	0.00
19 T	Methyl tert-butyl Ether	50.000	51.453	-2.9	106	0.00
20 T	Methylene Chloride	50.000	47.557	4.9	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.760	2.5	111	0.00
22 T	Diisopropyl ether	50.000	51.347	-2.7	106	0.00
23 T	Vinyl Acetate	250.000	262.691	-5.1	105	0.00
24 P	1,1-Dichloroethane	50.000	50.685	-1.4	111	0.00
25 T	2-Butanone	250.000	262.240	-4.9	111	0.01
26 T	2,2-Dichloropropane	50.000	46.538	6.9	106	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.111	-0.2	108	0.00
28 T	Bromochloromethane	50.000	48.611	2.8	105	0.00
29 T	Tetrahydrofuran	250.000	270.267	-8.1	112	0.00
30 C	Chloroform	50.000	50.597	-1.2#	108	0.00
31 T	Cyclohexane	50.000	50.512	-1.0	122	0.01
32 T	1,1,1-Trichloroethane	50.000	51.809	-3.6	119	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.835	-1.7	109	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	110	0.00
35 S	Dibromofluoromethane	50.000	49.967	0.1	109	0.00
36 T	1,1-Dichloropropene	50.000	48.814	2.4	118	0.00
37 T	Ethyl Acetate	50.000	52.084	-4.2	111	0.00
38 T	Carbon Tetrachloride	50.000	50.630	-1.3	120	0.00
39 T	Methylcyclohexane	50.000	49.203	1.6	123	0.00
40 TM	Benzene	50.000	49.339	1.3	109	0.00
41 T	Methacrylonitrile	50.000	53.042	-6.1	109	0.00
42 TM	1,2-Dichloroethane	50.000	49.733	0.5	106	0.00
43 T	Isopropyl Acetate	50.000	51.747	-3.5	109	0.00
44 TM	Trichloroethene	50.000	48.624	2.8	113	0.00
45 C	1,2-Dichloropropane	50.000	49.973	0.1#	108	0.00
46 T	Dibromomethane	50.000	48.572	2.9	105	0.00
47 T	Bromodichloromethane	50.000	51.101	-2.2	108	0.00
48 T	Methyl methacrylate	50.000	53.021	-6.0	109	0.00
49 T	1,4-Dioxane	1000.000	1090.779	-9.1	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122220\
 Data File : VX020316.D
 Acq On : 22 Dec 2020 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 23 00:36:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 18 13:09:49 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

50 S	Toluene-d8	50.000	49.811	0.4	111	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.431	-7.4	112	0.00
52 CM	Toluene	50.000	50.424	-0.8#	111	0.00
53 T	t-1,3-Dichloropropene	50.000	50.976	-2.0	105	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.682	-1.4	105	0.00
55 T	1,1,2-Trichloroethane	50.000	50.236	-0.5	107	0.00
56 T	Ethyl methacrylate	50.000	53.735	-7.5	109	0.00
57 T	1,3-Dichloropropane	50.000	50.473	-0.9	107	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	241.154	3.5	101	0.00
59 T	2-Hexanone	250.000	274.576	-9.8	114	0.00
60 T	Dibromochloromethane	50.000	52.002	-4.0	108	0.00
61 T	1,2-Dibromoethane	50.000	50.365	-0.7	106	0.00
62 S	4-Bromofluorobenzene	50.000	50.891	-1.8	114	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	48.805	2.4	115	0.00
65 PM	Chlorobenzene	50.000	49.662	0.7	108	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.700	-1.4	107	0.00
67 C	Ethyl Benzene	50.000	51.643	-3.3#	112	0.00
68 T	m/p-Xylenes	100.000	103.932	-3.9	111	0.00
69 T	o-Xylene	50.000	51.708	-3.4	108	0.00
70 T	Styrene	50.000	53.130	-6.3	108	0.00
71 P	Bromoform	50.000	54.795	-9.6	115	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	109	0.00
73 T	Isopropylbenzene	50.000	50.878	-1.8	116	0.00
74 T	N-aryl acetate	50.000	52.984	-6.0	110	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.528	-1.1	111	0.00
76 T	1,2,3-Trichloropropane	50.000	49.719	0.6	110	0.00
77 T	Bromobenzene	50.000	47.258	5.5	108	0.00
78 T	n-propylbenzene	50.000	50.514	-1.0	115	0.00
79 T	2-Chlorotoluene	50.000	49.274	1.5	110	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.880	-1.8	113	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.976	0.0	105	0.00
82 T	4-Chlorotoluene	50.000	49.521	1.0	110	0.00
83 T	tert-Butylbenzene	50.000	51.070	-2.1	116	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.137	-2.3	110	0.00
85 T	sec-Butylbenzene	50.000	51.427	-2.9	121	0.00
86 T	p-Isopropyltoluene	50.000	51.236	-2.5	116	0.00
87 T	1,3-Dichlorobenzene	50.000	48.065	3.9	107	0.00
88 T	1,4-Dichlorobenzene	50.000	46.643	6.7	107	0.00
89 T	n-Butylbenzene	50.000	51.077	-2.2	118	0.00
90 T	Hexachloroethane	50.000	50.862	-1.7	119	0.00
91 T	1,2-Dichlorobenzene	50.000	48.349	3.3	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.615	-3.2	110	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.712	2.6	110	0.00
94 T	Hexachlorobutadiene	50.000	50.701	-1.4	134	0.00
95 T	Naphthalene	50.000	53.072	-6.1	111	0.00
96 T	1,2,3-Trichlorobenzene	50.000	48.728	2.5	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122220\
Data File : VX020316.D
Acq On : 22 Dec 2020 20:33
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 23 00:36:18 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X121820W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 18 13:09:49 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6