

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
 Data File : VX039436.D
 Acq On : 22 Dec 2023 13:47
 Operator : JC/MD
 Sample : 05996-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-BR-04A-440-460-122123

Quant Time: Dec 23 06:24:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	104383	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	111738	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94146	49.234	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.460%
35) Dibromofluoromethane	5.379	113	73548	48.175	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.360%
50) Toluene-d8	8.647	98	283137	49.955	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.920%
62) 4-Bromofluorobenzene	11.079	95	124496	53.237	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.480%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	18678	13.915	ug/l	90
24) 1,1-Dichloroethane	3.611	63	3435	1.228	ug/l #	91
27) cis-1,2-Dichloroethene	4.489	96	14426	8.948	ug/l	90
42) 1,2-Dichloroethane	6.092	62	2826	1.024	ug/l	83
44) Trichloroethene	7.123	130	24773	14.405	ug/l	85
52) Toluene	8.720	92	27081	6.584	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
Data File : VX039436.D
Acq On : 22 Dec 2023 13:47
Operator : JC/MD
Sample : 05996-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-BR-04A-440-460-122123

Quant Time: Dec 23 06:24:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

