

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
 Data File : VX039437.D
 Acq On : 22 Dec 2023 14:10
 Operator : JC/MD
 Sample : 05966-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COLT3

Quant Time: Dec 23 06:24:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	103988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221740	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231986	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	94288	49.495	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.000%
35) Dibromofluoromethane	5.391	113	72902	47.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.180%
50) Toluene-d8	8.647	98	282318	49.641	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.280%
62) 4-Bromofluorobenzene	11.079	95	127456	54.317	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
Data File : VX039437.D
Acq On : 22 Dec 2023 14:10
Operator : JC/MD
Sample : 05966-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COLT3

Quant Time: Dec 23 06:24:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

