

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
 Data File : VX039438.D
 Acq On : 22 Dec 2023 14:32
 Operator : JC/MD
 Sample : 05966-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMA7

Quant Time: Dec 23 06:25:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	113493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92898	50.804	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.600%	
35) Dibromofluoromethane	5.385	113	70603	46.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.480%	
50) Toluene-d8	8.647	98	276749	49.348	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.700%	
62) 4-Bromofluorobenzene	11.079	95	123672	53.448	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.900%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
Data File : VX039438.D
Acq On : 22 Dec 2023 14:32
Operator : JC/MD
Sample : 05966-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMA7

Quant Time: Dec 23 06:25:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

