

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
 Data File : VX039439.D
 Acq On : 22 Dec 2023 14:55
 Operator : JC/MD
 Sample : 05966-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COLK8

Quant Time: Dec 23 06:25:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	103253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	217597	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227944	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114524	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93494	49.428	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.860%
35) Dibromofluoromethane	5.385	113	71340	47.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.920%
50) Toluene-d8	8.647	98	277518	49.726	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.460%
62) 4-Bromofluorobenzene	11.079	95	122027	52.994	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
Data File : VX039439.D
Acq On : 22 Dec 2023 14:55
Operator : JC/MD
Sample : 05966-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COLK8

Quant Time: Dec 23 06:25:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration

