

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
 Data File : VX039438.D
 Acq On : 22 Dec 2023 14:32
 Operator : JC/MD
 Sample : 05966-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMA7

Quant Time: Dec 23 06:25:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 01:28:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	99816	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218657	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226858	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	113493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92898	50.804	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.600%	
35) Dibromofluoromethane	5.385	113	70603	46.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.480%	
50) Toluene-d8	8.647	98	276749	49.348	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.700%	
62) 4-Bromofluorobenzene	11.079	95	123672	53.448	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.900%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
Data File : VX039438.D
Acq On : 22 Dec 2023 14:32
Operator : JC/MD
Sample : 05966-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0MA7

Quant Time: Dec 23 06:25:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 01:28:15 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.000 min
Lab File: VX039438.D
Acq: 22 Dec 2023 14:32




Instrument :
MSVOA_X
ClientSampleId :
COMA7

Tgt Ion:168 Resp: 99816
Ion Ratio Lower Upper
168 100
99 69.8 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 50.804 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.000 min
Lab File: VX039438.D
Acq: 22 Dec 2023 14:32

Tgt Ion: 65 Resp: 92898
Ion Ratio Lower Upper
65 100
67 52.5 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.000 min
Lab File: VX039438.D
Acq: 22 Dec 2023 14:32




Instrument :
MSVOA_X
ClientSampleId :
COMA7


Tgt Ion:114 Resp: 218657
Ion Ratio Lower Upper
114 100
63 22.6 0.0 45.0
88 16.8 0.0 36.6



#35
Dibromofluoromethane
Concen: 46.739 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX039438.D
Acq: 22 Dec 2023 14:32

Tgt Ion:113 Resp: 70603
Ion Ratio Lower Upper
113 100
111 105.0 82.9 124.3
192 15.7 13.3 19.9





#50

Toluene-d8

Concen: 49.348 ug/l

RT: 8.647 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX039438.D

Acq: 22 Dec 2023 14:32

Instrument :

MSVOA_X

ClientSampleId :





COMA7

Tgt Ion: 98 Resp: 276749

Ion Ratio Lower Upper

98 100

100 64.0 53.1 79.7



#62

4-Bromofluorobenzene

Concen: 53.448 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX039438.D

Acq: 22 Dec 2023 14:32




Tgt Ion: 95 Resp: 123672


Ion Ratio Lower Upper

95 100

174 63.1 0.0 135.6

176 59.9 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX039438.D
Acq: 22 Dec 2023 14:32

Instrument :
MSVOA_X
ClientSampleId :
COMA7

Tgt Ion:117 Resp: 226858
Ion Ratio Lower Upper
117 100
82 59.9 46.2 69.4
119 31.2 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.000 min
Lab File: VX039438.D
Acq: 22 Dec 2023 14:32

Tgt Ion:152 Resp: 113493
Ion Ratio Lower Upper
152 100
115 67.4 43.3 129.9
150 156.4 0.0 345.0

