



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX122223\
Data File : VX039430.D
Acq On : 22 Dec 2023 11:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X121323W.M
Title : SW846 8260
Last Update : Thu Dec 14 01:28:15 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	6770	PASS
75	95	30	60	56.4	17612	PASS
95	95	100	100	100.0	31213	PASS
96	95	5	9	6.4	2013	PASS
173	174	0.00	2	1.2	233	PASS
174	95	50	100	62.8	19606	PASS
175	174	5	9	7.0	1379	PASS
176	174	95	101	96.0	18824	PASS
177	176	5	9	6.6	1242	PASS